

Biosensor Förster Resonance Energy Transfer Detection by the Phasor Approach to Fluorescence Lifetime Imaging Microscopy

ELIZABETH HINDE,¹ MICHELLE A. DIGMAN,^{1,2} CHRISTOPHER WELCH,³ KLAUS M. HAHN,³
AND ENRICO GRATTON^{1,2*}

¹Laboratory for Fluorescence Dynamics, Department of Biomedical Engineering, University of California, Irvine, California

²Department of Development and Cell Biology, University of California, Irvine, California

³Department of Pharmacology, University of North Carolina, Chapel Hill, North Carolina

KEY WORDS fluorescence lifetime; phasor analysis; Förster resonance energy transfer; biosensor

ABSTRACT We present here the phasor approach to biosensor Förster resonance energy transfer (FRET) detection by fluorescence lifetime imaging microscopy (FLIM) and show that this method of data representation is robust towards biosensor design as well as the fluorescence artifacts inherent to the cellular environment. We demonstrate this property on a series of dual and single chain biosensors, which report the localization of Rac1 and RhoA activity, whilst performing concomitant ratiometric FRET analysis on the acquired FLIM data by the generalized polarization (GP) approach. We then evaluate and compare the ability of these two methods to quantitatively image biosensor FRET signal as a function of time and space. We find that with lifetime analysis in the phasor plot each molecular species is transformed into a two-dimensional coordinate system where independent mixtures of fluorophores can be distinguished from changes in lifetime due to FRET. This enables the fractional contribution of the free and bound state of a dual chain biosensor or the low and high FRET species of a single chain biosensor to be quantified in each pixel of an image. The physical properties intrinsic to each biosensor design are also accurately characterized by the phasor analysis; thus, this method could be used to inform biosensor optimization at the developmental stage. We believe that as biosensors become more sophisticated and are multiplexed with other fluorescent molecular tools, biosensor FRET detection by the phasor approach to FLIM will not only become imperative to their use but also their advancement. *Microsc. Res. Tech.*

75:271–281, 2012. © 2011 Wiley Periodicals, Inc.

INTRODUCTION

The development of fluorescent protein biosensors coupled with live-cell imaging has enabled the visualization and measurement of intracellular molecular dynamics with high spatiotemporal resolution (Gaits and Hahn, 2003; Hahn and Touthkine, 2002). Fluorescent biosensors designed to probe kinase, phosphatase and GTPase activity, second messenger dynamics, metabolites such as glucose, and receptor–effector coupling are just a few examples of the continuously expanding collection currently available (DiPilato and Zhang, 2010; Wallrabe and Periasamy, 2005). The great majority of biosensors based on fluorescent proteins use an inherent FRET (Förster resonance energy transfer) interaction to respond to the level of cellular activity being probed (Ibraheem and Campbell, 2010; Tsien, 1998). Given that these molecular tools are designed to report on the spatial localization, temporal coordination, and biochemical concentrations of specific signaling events, detection and quantitation of FRET as a function of time and space in cells is a matter of great interest.

FRET is a phenomenon where a donor fluorophore in an excited electronic state transfers its excitation energy to a nearby acceptor fluorophore via nonradia-

tive dipole-dipole interaction (Clegg, 1995; Förster, 1965). Distance and orientation determine whether a donor-acceptor fluorophore pair can undergo FRET, and this is the key to differentiating the on and off state (dynamic range) of a FRET biosensor (Tsien, 1998). When a FRET interaction does occur, the donor fluorescence intensity and lifetime are quenched and the acceptor fluorescence intensity is enhanced, as well as shifted in lifetime (Lakowicz and Masters, 2008). In theory, any one of these changes can be used to detect the spatiotemporal dynamics of a FRET biosensor and thereby generate a FRET image. In practice, the technical challenge is to quantify and differentiate a FRET signal from the several other sources of fluorescence

Additional Supporting Information may be found in the online version of this article.

*Correspondence to: Enrico Gratton, Laboratory for Fluorescence Dynamics, Department of Biomedical Engineering, University of California, Irvine, CA, United States. E-mail: egratton22@yahoo.com

Received 11 June 2011; accepted in revised form 12 June 2011

Contract grant sponsor: NIH-P41; Contract grant number: P41-RRO3155; Contract grant sponsor: NIH; Contract grant number: P50-GM076516; Contract grant sponsor: The Cell Migration Consortium; Contract grant numbers: U54 GM064346, GM057464; Contract grant sponsor: NIH-T32; Contract grant number: GM008719; Contract grant sponsor: NIH-F30; Contract grant number: F30HL094020-02

DOI 10.1002/jemt.21054

Published online 19 August 2011 in Wiley Online Library (wileyonlinelibrary.com).

that are present in live cells (Periasamy et al., 2008; Wallrabe and Periasamy, 2005).

A FRET biosensor can be constructed using either a dual chain or single chain design (Hodgson et al., 2010; Tsien, 1998). A dual chain design places the donor and acceptor on two separate molecules (chains), which only bind upon activation. This design is advantageous in terms of dynamic range, in that there is no baseline interaction between the two-biosensor chains. However, by employing two chains, there is a propensity for the donor and acceptor fluorophores to distribute unequally throughout the cell and this causes artifact for intensity based FRET detection (Hodgson et al., 2010). A single chain design places the donor and acceptor on the same molecule and activation of the target causes a conformational change that alters the proximity and/or orientation of the fluorophores (Hodgson et al., 2008, 2010). This design is advantageous for ease of image analysis. However, by connecting the donor and acceptor, they are in close enough proximity to cause an inherent residual FRET signal that does not originate from the cellular activity being probed.

We present here the phasor approach to biosensor FRET detection by fluorescence lifetime imaging microscopy (FLIM) and show that this method of data representation is robust towards biosensor design and the fluorescent artifacts inherent to a cellular environment. We demonstrate this on a series of dual and single chain GTPase biosensors, which report the localization of Rac1 and RhoA activity (Hodgson et al., 2008; 2010; Kraynov et al., 2000; Machacek et al., 2009; Pertz and Hahn, 2004). We perform concomitant ratiometric FRET analysis on the acquired FLIM data by the generalized polarization (GP) approach (Parasassi et al., 1991), demonstrating the advantages of the phasor approach.

Rac1 is known to promote membrane protrusion at the leading edge, while RhoA regulates contractility in the cell body and at adhesions throughout the cell (Burridge and Wennerberg, 2004). Recent studies using the biosensors employed here indicate that Rac and Rho interact at the leading edge to regulate protrusion, with RhoA activation being synchronous with edge advancement and Rac1 shifted 2 μm behind (Machacek et al., 2009; Pertz et al., 2006). With these spatiotemporal dynamics in mind, we evaluate the ability of each method to quantitatively image a FRET signal as a function of time and space.

Determination of FRET Using Intensity Based Ratiometric Analysis and the GP Function

Quantitation of the FRET signal from a biosensor requires knowledge of the relative concentrations of the different states of the biosensor: the free and bound state of a dual chain design and the low FRET (LF) and high FRET (HF) state of a single chain design (Hodgson et al., 2010). In either case, the various forms of the biosensor are a linear combination of the spectra of the donor and the acceptor fluorophores; thus, an intensity based ratiometric method of analysis should be sufficient to derive the relative concentration of the two species after calibration for instrumental artifacts (e.g., spectral bleed through). In the cellular environment this is not the case, since not all of the biosensor

expressed by the cell is optically active. In all cases of fluorescent protein expression there is a fraction of the protein that does not completely mature (Llères et al., 2001). This leads to biosensor fluorescence at donor and/or acceptor wavelengths that are not responsive to the physiological state being measured. Furthermore the two fluorescent proteins (donor and acceptor) invariably experience photobleaching at different rates (Hodgson et al., 2006) and this also results in a population of unresponsive donor only and acceptor only FRET pairs, which accumulate over the duration of the experiment (Figs. 1A and 1B).

Determination of the individual concentrations of the two states of the biosensor based on intensity is thus virtually impossible; only relative changes in the population of the on versus off state can be ascertained by a ratiometric method of analysis (Fig. 1C). One way to determine this relative change is to calculate the normalized FRET ratio of each pixel in an image by use of the GP function [defined in Eqs. (10)–(12) of the materials and methods section]. The GP function has been used extensively to quantify the change in spectrum of Laurdan in response to membrane fluidity (Parasassi et al., 1991; Yu et al., 1996). The GP function can also be used to measure the change in emission spectrum due to FRET and has the advantageous property of transforming the concentration of each species into becoming additive, since the denominator of the GP is proportional to the total fluorescence intensity (Jameson et al., 1984). This enables a graphical representation, where if we know the GP of the on and off state of a biosensor, then the relative GP of an unknown mixture can be found from the linear combination of the two known GPs (Fig. 1D). It is simply the alignment of the donor and acceptor only spectra with the on and off state of a biosensor that prevents the determination of the individual species by spectral unmixing (Figs. 1E and 1F).

Determination of FRET Using Quenching of the Donor Fluorescence and the Phasor Approach to FLIM

Determination of FRET in cells at a quantitative level is thought to be best achieved using the quenching of donor lifetime, since this parameter is changed to a value that can be resolved from the on and off state of a dual or single chain biosensor (Bastiaens and Squire, 1999; Suhling et al., 2005; Wallrabe and Periasamy, 2005). FLIM measures the average lifetime of the donor fluorescence emission in the presence of the acceptor in each pixel of an image and with prior knowledge of the donor lifetime in the absence of the acceptor, assigns those pixels producing a quenched lifetime to the FRET localization. There are different approaches available to detect FRET by FLIM, distinguished by image acquisition in either the time or the frequency domain (Chandler et al., 2006; van Munster EB, 2005; Wallrabe and Periasamy, 2005). FLIM data is primarily acquired in the time domain with analysis proceeding by fitting the average fluorescence decay in each pixel using an exponential model. However, this mode of analysis presents a formidable computational problem when multiple fluorescent species are present as is always the case in any FRET experiment.




Fig. 1. Intensity based biosensor FRET detection of a dual and single chain biosensor. **A:** The fluorescent species encountered in dual chain biosensor FRET detection: donor chain, acceptor chain, and the donor and acceptor chain bound. **B:** The fluorescent species encountered in single chain biosensor detection: optically inactive biosensor, low FRET biosensor, and high FRET biosensor. **C:** Emission spectra of the various fluorescent species present in a FRET experiment that must be linearly unmixed when performing ratiometric analysis of a single chain or dual chain biosensor. **D:** Graphical representation of the additive property of GP transformation. If we know the GP of the on and off state of the biosensor, then the fractional contribution of the

GP of an unknown mixture can be found by the linear combination of GP1 and GP2. The fractional contribution of a third species with a different spectrum (e.g., cell autofluorescence) can also be determined in each pixel with detection at a third wavelength. **E** and **F:** The alignment of the 3 or 4 GP values in the dual or single chain FRET experiment, respectively, prevents the determination of the contribution of the individual species by spectral unmixing; unlike the situation illustrated with the phasor approach (Fig. 2). [Color figure can be viewed in the online issue, which is available at wileyonlinelibrary.com.]

Analysis of FLIM data in the time domain can be simplified by use of the phasor approach, a vector representation conventionally employed to analyze data collected in the frequency domain (Clayton et al., 2004; Digman et al., 2008; Gratton et al., 1984; Jameson et al., 1984; Lakowicz et al., 1984). The phasor approach provides a global view of the fluorescence decay in an image by transforming the histogram of time delays in each pixel into a phasor. The sine-cosine transforms of each phasor are plotted in a two dimensional space termed the universal plot, and each phasor position is characteristic of a particular molecular species and its local environment (Colyer et al., 2008; Digman et al., 2008). In the phasor space, you can distinguish a mixture of independent molecular species (which form a linear trajectory) from a change in lifetime due to FRET (which forms a curved trajectory) without having to resolve the decay at each pixel into the individual exponential components (Digman et al., 2008). Figure 2 illustrates this concept and shows that irrespective of biosensor design, analysis of a FRET signal

in the phasor plot enables the various fluorescent species present in a FRET experiment to be resolved. The striking difference of the phasor representation compared with the intensity ratio in Figure 1 is that FRET changes the lifetime so that a new phasor position is produced that is not aligned with the on and off state of the biosensor.

MATERIALS AND METHODS

Cell Culture and Treatments

COS7 cells were grown in high glucose medium from Invitrogen, supplemented with 10% of Fetal Bovine Serum, 5 mL of Pen-Strep and HEPES at 37°C and in 5% CO₂. Freshly split cells were plated onto 35-mm glass bottom dishes coated with fibronectin and then after 24 hours, transiently transfected with the following series of GTPase biosensors sourced from the Hahn Laboratory, University of North Carolina: (1) Rac1 dual chain biosensor (CyPet-Rac1 and YPet-PBD); (2) Rac1 dominant negative dual chain biosensor




Fig. 2. Analysis of the FRET signal from a dual chain and single chain biosensor in the phasor plot. **A:** Dual chain design: the phasor location of the donor-only species and autofluorescence form the linear combination from which a phasor distribution due to FRET will originate. By superimposing a FRET trajectory [defined by Eq. (9) in the Materials and Methods section] over the most quenched lifetime we can determine the efficiency of the FRET state of the biosensor, and from the line that connects this FRET state back to the donor, quantitate the population of donor undergoing FRET in each pixel. **B:** Single chain design: the phasor location of the donor-only species

is no longer the off state of the biosensor. Instead, it is a low FRET state due to a residual FRET signal inherent to the single chain biosensor design. In this scenario the low FRET state, which is shifted toward a shorter lifetime along the FRET trajectory from the donor phasor is the off state and then further shifted to the most quenched lifetime along the FRET trajectory is the high FRET state. The population of molecules undergoing FRET can be quantified from the line that connects these two phasor positions. [Color figure can be viewed in the online issue, which is available at www.interscience.wiley.com.]

(CyPet-T17N-Rac1 and YPet-PBD); (3) Rac1 constitutively active dual chain biosensor of (CyPet-Q61L-Rac1 and YPet-PBD); (4) RhoA single chain biosensor (RBD-Citrine-1L-ECFP-RhoA); (5) RhoA constitutively membrane bound biosensor (RBD-Citrine-1L-ECFP-RhoA-kRas). The plated cells were then left for a further 24 hours at 37°C/5% CO₂ and then serum starved in unsupplemented high glucose medium. The Rac biosensors tested were stimulated with 50 ng mL⁻¹ of epidermal growth factor (EGF), Sigma Aldrich. The Rho biosensors tested were stimulated with 2.3 μM of lysophosphatidic acid (LPA), Cayman.

Microscope

FLIM and ratiometric data were acquired concomitantly with the Zeiss LSM710 META laser scanning microscope, coupled to a 2-photon Ti:Sapphire laser (Spectra-Physics Mai Tai, Newport Beach) producing 80 fs pulses at a repetition of 80 MHz, and a ISS A320 FastFLIM box to acquire the lifetime data. A 63× water immersion objective 1.2 N.A. (Zeiss, Germany) was used for all experiments. The donor fluorophore of each biosensor was excited at 800 nm with the 2-photon laser: this wavelength caused negligible direct excitation of the acceptor fluorophore. A SP 760 nm dichroic filter was used to separate the fluorescence signal from the laser light. The fluorescence signal was directed through a 509 LP CFP/YFP filter, and the donor and acceptor signal split between two photomultiplier detectors (H7422P-40 of Hamamatsu), with the following bandwidth filters in front of each: CFP 470/22 and YFP 542/27, respectively. For image acquisition, the pixel frame size was set to 256 × 256 and the pixel dwell time to 25.61 μs/pixel. The average laser power at the sample was maintained at the mW level. The FLIM and ratiometric data were acquired and processed by the SimFCS software developed at the Laboratory for Fluorescence Dynamics (www.lfd.uci.edu). Calibration of the system and pha-

sor plot space was performed by measuring fluorescein (pH 9.0), which has a known single exponential lifetime of 4.04 ns.

Data Analysis

Phasor Transformation. The phasor transformation and data analysis were performed using the SimFCS software available at the www.lfd.uci.edu. Every pixel of the FLIM image is transformed into one pixel in the phasor plot. The *s* and *g* coordinates in the phasor plot, which correspond to a given decay *I*(*t*) are defined by the expressions by:

$$g_{ij}(\omega) = \frac{\int_0^{\infty} I_{ij}(t) \cos(\omega t) dt}{\int_0^{\infty} I_{ij}(t) dt}, \quad (1)$$

$$S_{ij}(\omega) = \frac{\int_0^{\infty} I_{ij}(t) \sin(\omega t) dt}{\int_0^{\infty} I_{ij}(t) dt}, \quad (2)$$

where ω is the laser repetition angular frequency, and the indexes *i* and *j* identify a pixel of the image. If the data are measured in the frequency domain then:

$$g_{ij}(\omega) = m_{ij} \cos(\phi_{ij}), \quad (3)$$

$$S_{ij}(\omega) = m_{ij} \sin(\phi_{ij}), \quad (4)$$

where m_{ij} and ϕ_{ij} are the modulation and the phase of the emission with respect to the excitation. If the decay is single exponential $I(t) = Ae^{-t/\tau}$ the coordinates of the phasor are given by:

$$g_{ij}(\omega) = \frac{1}{1 + (\omega\tau)^2}, \quad (5)$$

$$S_{ij}(\omega) = \frac{\omega\tau}{1 + (\omega\tau)^2}. \quad (6)$$

In the case, where a pixel ij has the contribution of several exponential components, the coordinates g and s of the phasor are given by:

$$g_{ij}(\omega) = \Sigma_k \frac{f_k}{1 + (\omega\tau_k)^2}, \quad (7)$$

$$S_{ij}(\omega) = \Sigma_k \frac{f_k \omega\tau_k}{1 + (\omega\tau_k)^2}, \quad (8)$$

where f_k is the intensity weighted fractional contribution of the component with lifetime τ_k . According to expressions 5 and 6 for the coordinate of a phasor for a single exponential decay

$$S_{ij}^2 + \left(g_{ij} - \frac{1}{2}\right)^2 = \frac{1}{4}$$

which implies that all single exponential components are represented by a semicircle of center $(1/2, 0)$ and radius $1/2$ in the phasor plot. We name this the “universal circle.” Along this semicircle a phasor corresponding to a very short lifetime (small phase angle) is close to the point $(1, 0)$, where as a phasor corresponding to a very long lifetime will be close to the $(0, 0)$ point.

Resolution of Two Components. To resolve the fractional contribution of two phasor components we graphically solve Eqs. (7) and (8) in the phasor plot. All phasors corresponding to the combination of the component phasors are along a segment where the segment extremes correspond to the phasors of the two isolated components. Generally, the locations of the two component phasors are known or are derived from the experimental points of systems in which all possible combinations are realized. The segment is drawn on the phasor plot and the operator selects points along the segment to match the experimental points. The fractional contribution of the two phasors at the cursor location is displayed on the screen. Using the reciprocal property of the cursor, the points in the image corresponding to that particular solution are highlighted.

FRET Efficiency Calculation. The trajectory is calculated according to the classical definition of FRET efficiency:

$$E = 1 - \frac{\tau}{\tau_D} \quad (9)$$

The phasor of the donor in the absence of the acceptor is obtained from an independent preparation in which the acceptor is absent. The phasor corresponding to the quenched donor is calculated according to the quenching Eq. (9). The realizations of all possible phasors that are quenched with different efficiencies describe a curved trajectory in the phasor plot. The

experimental position of the phasor of a given pixel along the trajectory determines the amount of quenching and therefore the FRET efficiency. The contributions of the background and of the donor without acceptor are evaluated using the rule of the linear combination [Eqs. (7) and (8)], with the background phasor and the donor unquenched determined independently. As depicted in Figure 2 the operator moves the cursor along the curved trajectory to the point where there is a cluster of phasors. Using the reciprocal property of the cursor, the pixels in the image corresponding to a given FRET efficiency are highlighted.

Normalized Ratio or GP. For the intensity based ratiometric analysis, we employ the normalized ratio approach, which has the property of transforming the concentrations of the various species additively. The normalized ratio function calculates the normalized FRET ratio, which enables visualization of the changes between the two FRET states. The GP function can be used when there is a change in the emission spectrum due to FRET and automatically compensates for bleed through. The GP function is related to the normal ratiometric methods of analysis and defined as follows:

$$R = \frac{I_{\text{FRET}}}{I_{\text{Donor}}}, \quad (10)$$

$$\text{GP} = \frac{R - 1}{R + 1}, \quad (11)$$

$$\text{GP} = \frac{I_{\text{FRET}} - I_{\text{Donor}}}{I_{\text{FRET}} + I_{\text{Donor}}}, \quad (12)$$

where I_1 and I_2 are the fluorescence intensities at the spectral band pass of fluorophores 1 and 2, respectively. Mathematically, the GP function is always bound between -1 and $+1$. The GP function depends on the choice of the band pass filters and given two band pass filters, each fluorescent substance has a specific value of the GP. When the fluorescence intensity is measured in a “spectrally corrected” fluorometer, the correct GP value can be calculated directly. If we know the corrected GP value for a molecular species, for example for fluorescein or any other commonly available dye, we can calibrate an unknown system (generally the microscope) using the following expression and correction factor g :

$$\text{GP} = \frac{I_{\text{FRET}} - gI_{\text{Donor}}}{I_{\text{FRET}} + gI_{\text{Donor}}}, \quad (13)$$

where g is the correction factor calculated from the following expression:

$$g = \left(\frac{1 - \text{GP}_c}{1 + \text{GP}_c} \right) \left(\frac{1 + \text{GP}_r}{1 - \text{GP}_r} \right), \quad (14)$$

where GP_c is the known value of the GP for the calibration compound and GP_r is the value of the raw GP of the same substance measured in the particular instrument that needs to be calibrated. If the denominator of the GP function is proportional to the total fluorescence

intensity then the GP function of a mixture of two fluorophores is equal to the intensity-weighted average of the GP of the two separate species. This property of the GP function is referred to as the “additive property of the GP” and it will be used throughout. Using the GP, the fraction α of biosensor in the bound form can be directly calculated from the following equation:

$$\alpha = \frac{GP - kGP_{\text{free}}}{GP_{\text{bound}} - kGP_{\text{free}}}, \quad (15)$$

where k is the ratio of the quantum yield of the biosensor in the two forms. As mentioned in the introduction, when working with cells, there are a series of problems with the direct application of ratiometric formula like the GP function. First, it is unclear whether the spectra of the two forms of the biosensor (bound and free) remain the same in the cell environment. Second, the measured GP contains the contribution arising from autofluorescence. Figure 1 illustrates how to use the GP approach: in the vertical axis, we plot the GP of the two species, for example high FRET and Low FRET, and for every GP on the y -axis, we obtain the fractional contribution of the two species using the linear combination rule. However, for every system with more than two components, the derivation of the fractional intensity given the GP cannot be obtained using only two emission wavelengths, because the value of the GP of the individual components must first be combined with the GP value of the third species, which requires the knowledge of the relative fractional intensity of the third component.

RESULTS

Dual Chain Biosensor FRET Detection

The Rac1 dual chain biosensor tested employs a blue-green fluorescent protein CyPet (donor) and a yellow fluorescent protein YPet (acceptor) as the FRET pair. On the donor chain, we have the targeted protein Rac1 and on the acceptor chain we have the affinity reagent PBD [binding domain derived from p21-activated kinase1 (PAK1)]. The affinity reagent PBD interacts only with the activated state of Rac1, which is induced here by EGF stimulation. Upon activation, the two fluorescent proteins are brought into an orientation of sufficient spatial proximity to generate FRET. We measure concomitantly the lifetime and intensity image of the Rac1 biosensor in the donor and FRET channel following the experimental set up specified in the Materials and Methods section. This enables a direct comparison of FRET localization by lifetime analysis with FRET localization by intensity analysis on the same data set.

For the FLIM analysis we need only analyze the donor channel and determine the spatial distribution of FRET based on the degree of quenching of the donor lifetime. The precision of the phasor location depends on the number of counts in each pixel and so each measurement was integrated until there was a minimum of 100 counts per pixel (~ 10 frames). In general, we acquired a measurement before EGF stimulation and then 12 measurements after EGF stimulation each measurement lasting about 10 s. FLIM analysis of FRET by the phasor approach requires prior knowl-

edge of the donor phasor in the absence of acceptor. We therefore first measured a COS7 cell transfected with only the donor chain CyPet-Rac1 before and after EGF stimulation (Fig. 3A), to locate the phasor of the donor and to check for potential artifacts due to addition of EGF. As seen in the phasor plot of CyPet-Rac1 (Fig. 3B), addition of EGF does not cause a shift in the donor phasor (selected within the cyan cursor), which has an average lifetime of 2.61 ns. By use of the reciprocal property of the phasor cursor, we highlight all the pixels in the intensity images of the cell that contain this average lifetime (Fig. 3C).

We then measured a COS7 cell transfected with both donor (CyPet-Rac1) and acceptor (YPet-PBD) before and after EGF stimulation (Fig. 3D); the phasor distribution for this experiment is shown in Figure 3E. We first selected a region in this phasor plot that included only pixels in the background, and marked this as the phasor location for cellular auto-fluorescence (black cursor). A cyan cursor was then placed in the unquenched donor position (2.61 ns) and the FRET state of the dual chain biosensor extrapolated by superimposing a theoretical trajectory of all possible donor phasors [defined by Eq. (9) of the Materials and Methods section] over the experimentally measured distribution. This trajectory is depicted in Figure 3F and as can be seen the efficiency of the FRET state (where 100% donor is quenched) is determined as 30%. The linear combination of phasor clusters between the donor phasor and FRET state of the biosensor represent the varying contributions of donor fluorescence and quenched donor fluorescence in any one pixel. By moving the phasor cursor along the straight line drawn between these two terminal phasor locations (Fig. 3F), we can calculate the exact fractional contribution of quenched donor fluorescence for each pixel highlighted.

The spatial distribution of FRET detected by FLIM in the cell over time is represented in Figure 3G by pseudo coloring the cell images with a palette defined by the series of cursors, which extend from the donor lifetime (inside the cyan circle) toward the FRET state. In this color scheme, pixels highlighted in blue correspond to 0% quenched donor, green to 12% quenched donor, yellow to 24% quenched donor, and red to 58% quenched donor. As can be seen in the FLIM images of Figure 3G before EGF stimulation the entire cell is blue, which corresponds to the free state of the biosensor. After addition of EGF the population of donor quenched in the pixels located toward the cell perimeter increases with time (see Supporting Information Fig. S1A for FLIM analysis of this region of interest). This indicates an increase in the number of molecules undergoing FRET in those pixel locations and therefore an increase in Rac1 activity. We next calculated the normalized FRET ratio in each pixel of the images in Figure 3A using the GP function and compared the localization of FRET detected by this method to the FRET localization we observed with the FLIM analysis.

The GP function calculates the difference in intensity between the FRET and donor fluorescence intensity images, which are normalized to the total fluorescence intensity; mathematically this limits a GP value between +1 and -1. As seen in Figure 3H before




Fig. 3. FRET detection of a Rac1 dual chain biosensor. **A:** CyPet-Rac1 before and after EGF stimulation (donor). **B:** Phasor plot of CyPet-Rac1 experiment: cluster selected within the cyan cursor corresponds to an average lifetime of 2.61 ns. **C:** Painted FLIM image of CyPet-Rac1. **D:** Intensity images of COS7 cell transfected with both donor (CyPet-Rac1) and acceptor (YPet-PBD) before and after

EGF stimulation (donor channel). **E:** Phasor plot of CyPet-Rac1 and YPet-PBD experiment. **F:** FRET efficiency calculator of CyPet-Rac1 and YPet-PBD experiment. **G:** Painted FLIM image of this CyPet-Rac1 and YPet-PBD experiment. **H:** Average GP of CyPet-Rac1 and YPet-PBD before and after EGF stimulation. **I:** Painted GP images of the CyPet-Rac1 and YPet-PBD experiment.

EGF stimulation the average GP value is -0.2 (GP of the free state of this specific biosensor). With time after EGF stimulation, the average GP value increases to a maximum of 0.5 (GP of the bound state of the biosensor). The spatial localization of each GP distribution and thus relative level of FRET activity is shown in Figure 3I. The derived GP images indicate that the perinuclear region of the cell is in an activated state over the duration of the experiment (even before EGF stimulation), and the lower right hand periphery of the cell experiences an incremental increase in FRET with time (see Supporting Information Fig. S1A for GP analysis of this region of interest). This localization of biosensor activity is markedly different from that shown by the FLIM analysis. This is because intensity based analysis is sensitive to large differences in the sub cellular concentration of biosensor components, here causing the low concentration of donor fluorophore in the perinuclear region (Fig. 3D) to be erroneously read

as quenching of donor fluorescence due to FRET. Control studies using the inactive (CyPet-T17N-Rac1) and constitutively active (CyPet-Q61L-Rac1) donor chain with the acceptor chain (YPet-PBD) showed only uniform low or high FRET respectively in both FLIM and GP analysis (Supporting Information Fig. S2).

Single Chain Biosensor FRET Detection

Here, we measure the FRET activity of two single chain RhoA biosensors (each designed for different sub-cellular localization) using the phasor and GP methods of analysis. We compare the results obtained to dual chain biosensor FRET detection. The first single chain RhoA biosensor tested employs ECFP (donor) and Citrine (acceptor), built into a chain that connects RhoA (the targeted protein) with RBD (a small RhoA binding domain derived from Rhotekin, the affinity reagent), leaving the C terminus free so that the protein is subject to normal regulatory pathways that control




Fig. 4. FRET detection of a single chain RhoA biosensor, with normal subcellular localization. **A:** Intensity images of COS7 cell transfected with (RBD-Citrine)-1L-(ECFP-RhoA) before and after LPA stimulation (donor channel). **B:** Phasor plot of (RBD-Citrine)-1L-

(ECFP-RhoA) experiment. **C:** FRET efficiency calculator of (RBD-Citrine)-1L-(ECFP-RhoA) experiment. **D:** Painted FLIM image of this (RBD-Citrine)-1L-(ECFP-RhoA) experiment. **E:** Painted GP images of an (RBD-Citrine)-1L-(ECFP-RhoA) experiment.

localization (i.e., binding of Guanine Nucleotide Dissociation Inhibitors, GDI). The rho-kinase fragment binds only to the activated state of RhoA, which is induced here by LPA stimulation. Upon activation, the distance and orientation between the ECFP and Citrine is altered such that the FRET efficiency of the biosensor is enhanced. We measure concomitantly the lifetime and intensity image of the RhoA biosensor in the donor and FRET channels, following the experimental set up specified under the Materials and Methods section and then compare the FRET localization obtained by the two analyses.

Figure 4A, depicts the donor intensity channel of a COS7 cell transfected with the RhoA single chain biosensor before and after LPA stimulation. The phasor distribution of this experiment is depicted in Figure 4B and the unquenched donor phasor position is marked within the cyan circle. As can be seen the

phasor distribution of this experiment does not extend from the donor phasor but instead from an origin that is right shifted to a shorter lifetime along the superimposed FRET trajectory depicted in Figure 4C. This corresponds to the low FRET state of the biosensor and is calculated to have a FRET efficiency of 4% as consequence of the residual FRET interaction inherent to the single chain design. Further shifted to the most quenched lifetime along the FRET trajectory is the high FRET state of this biosensor, which is calculated to have an efficiency of 35%. Thus, this RhoA single chain design enables a higher FRET efficiency than the coefficient derived by FLIM for the dual chain Rac1 biosensor, but at the expense of sensitivity. The residual FRET signal that is present even when the RhoA is in the off state decreases the dynamic range.




Fig. 5. FRET detection of a RhoA-kRas single chain biosensor, with constitutive membrane localization. **A:** Intensity images of COS7 cell transfected with (RBD-Citrine)-1L-(ECFP-RhoA-kRas) before and after LPA stimulation (donor channel). **B:** Phasor plot of (RBD-Citrine)-1L-(ECFP-RhoA-kRas) experiment. **C:** FRET efficiency

calculator of (RBD-Citrine)-1L-(ECFP-RhoA-kRas) experiment. **D:** Painted FLIM image of this (RBD-Citrine)-1L-(ECFP-RhoA-kRas) experiment. **E:** Painted GP images of (RBD-Citrine)-1L-(ECFP-RhoA-kRas) experiment.

The line that connects the high FRET phasor to the low FRET state gives the fractional population of biosensor active, above the baseline FRET signal in any one pixel. As with the dual chain biosensor analysis we pseudo color the cell images with a palette defined by the series of colored cursors, which extend from the low FRET state (inside the blue circle) toward the high FRET state (red circle). In this color scheme pixels highlighted in blue correspond to 0% high FRET, green to 22% high FRET, yellow to 38% high FRET, and red to 67% high FRET. The pseudo colored FLIM images show the pixels at the leading edge (top right hand corner) and within the back protrusions of the cell (bottom left hand corner), to show an increase in the population of high FRET state with time after LPA addition (see Supporting Information Fig. S1B for FLIM analysis of this region of

interest). This indicates an incremental increase in the number of molecules undergoing FRET in those pixel locations, which agrees spatially with where the literature reports RhoA activity to occur upon cell migration (Machacek et al., 2009). The GP analysis of this experiment (Fig. 4E) reveals analogous FRET localization to that obtained with the FLIM analysis. The agreement of the data derived from the two techniques relates to the fact that with a single chain design the two components of the biosensor are ensured to distribute equally throughout the cell.

The second single chain RhoA biosensor tested places the fluorophores in the same positions as that described for the first, but the C terminus is now modified by attachment of a lipid motif from K-Ras in place of the wild-type RhoA lipid modification motif. This modification causes constitutive membrane localization

through prevention of RhoA regulation by GDI, which regulate reversible membrane localization. Figure 5A depicts the donor intensity channel of a COS7 cell transfected with the RhoA membrane bound single chain biosensor, before and after LPA stimulation. The phasor distribution of this experiment is depicted in Figure 5B, and the unquenched donor phasor position is marked within the cyan circle. The phasor distribution of this experiment extends from the right shifted low FRET state of the biosensor (calculated to have a FRET efficiency of 7%) toward the high FRET state (calculated to have an efficiency of 39%). However, upon closer inspection of the phasor distribution along the line that connects the low FRET to high FRET phasor, we see that the baseline population of high FRET biosensor present is not equal to 0% (blue cursor). In any one pixel, we have 9% (green cursor) to 74% (red cursor) active biosensor. When we pseudo color the cell images with the palette defined by the cursors in Figure 5C, we see that the pixels along a broad section of the cell membrane experience FRET before and after LPA stimulation (Fig. 5D). The GP analysis of this experiment (Fig. 5E) reflects this baseline FRET interaction and active biosensor localization (see Supporting Information Fig. S1C for FLIM and GP analysis of this region of interest).

DISCUSSION

Detection of FRET by an intensity-based ratiometric method of analysis is advantageous for several reasons: it can be performed on almost every fluorescence microscope and has a good signal to noise ratio, which enables the detection of rapid changes in biochemical concentrations and spatial dynamics. The fundamental problem of ratiometric analysis is that in practice it is very sensitive to the fluorescence artifacts inherent in both a cellular environment and in biosensor designs. Linear un-mixing cannot accurately separate the non-responsive additional fluorescent species (i.e., which arise from differential photobleaching or incomplete protein expression) from the on and off FRET states of the optically active biosensor since these species have the same spectra. For example as demonstrated upon GP analysis of the dual chain Rac1 biosensor, the propensity of the donor and acceptor chains to distribute unequally resulted in the perinuclear region of the cell exhibiting a FRET signal that originated from the comparatively low expression level of donor in this zone. Intensity based analysis of a biosensor FRET signal is thus vulnerable to reporting a spatial localization of FRET that is not indicative of the actual biosensor activity, and is unable to determine the exact biochemical concentrations involved in the cellular event being probed.

Given that biosensors are designed to report on such molecular detail, biosensor FRET detection requires a method that is more quantitative, and for this reason fluorescence lifetime imaging (FLIM) is considered advantageous. We show here that FRET detection by the phasor approach to FLIM is able to unequivocally determine the local concentration of the free and bound state of a dual chain biosensor, the low and high FRET species of a single chain biosensor in each pixel of an

image, and thus provide a quantitative map of the biosensor activity as a function of time. The robustness of the phasor method is demonstrated by the fact that irrespective of a dual or single chain design, the map of Rac1 and RhoA activity derived is in agreement with the localization reported in the literature upon growth factor stimulation. We also show that the physical properties intrinsic to each biosensor design can be accurately characterized by phasor analysis and thus inform biosensor optimization at the developmental stage. For example, the dynamic range of the Rac1 dual chain biosensor is higher than that of the two RhoA single chain designs, which exhibit residual FRET before activation; preventing detection of low level protein activity.

Thus, in conclusion, we believe that the phasor approach to FLIM is an essential tool not only for the experimental use of FRET biosensors but also for their advancement. It is a reliable method of FRET detection that can accurately determine *in vivo* the spatiotemporal dynamics of single or dual chain biosensors independent of other sources of fluorescence. We believe that as biosensors become more sophisticated and are multiplexed with other fluorescent molecular tools, the ability of the phasor approach to transform each molecular species into a two dimensional coordinate system, which distinguishes independent mixtures from changes in lifetime due to FRET, without having to resolve the decay at each pixel into the individual exponential components, will be imperative.

ACKNOWLEDGMENTS

The authors thank Milka Titin for cultivating and transfecting the cells.

REFERENCES

- Bastiaens PIH, Squire A. 1999. Fluorescence lifetime imaging microscopy: Spatial resolution of biochemical processes in the cell. *Trends Cell Biol* 9:48–52.
- Burridge K, Wennerberg K. 2004. Rho and Rac Take Center Stage. *Cell* 116:167–179.
- Chandler D, Majumdar Z, Heiss G, Clegg R. 2006. Ruby crystal for demonstrating time- and frequency-domain methods of fluorescence lifetime measurements. *J Fluoresc* 16:793–807.
- Clayton AHA, Hanley QS, Verveer PJ. 2004. Graphical representation and multicomponent analysis of single-frequency fluorescence lifetime imaging microscopy data. *J Microsc* 213:1–5.
- Clegg RM. 1995. Fluorescence resonance energy transfer. *Curr Opin Biotechnol* 6:103–110.
- Colyer RA, Lee C, Gratton E. 2008. A novel fluorescence lifetime imaging system that optimizes photon efficiency. *Microsc Res Tech* 71:201–213.
- Digman MA, Caiolfa VR, Zamai M, Gratton E. 2008. The phasor approach to fluorescence lifetime imaging analysis. *Biophys J* 94:L14–L16.
- DiPilato LM, Zhang J. 2010. Fluorescent protein-based biosensors: Resolving spatiotemporal dynamics of signaling. *Curr Opin Chem Biol* 14:37–42.
- Förster T. 1965. Delocalized excitation and excitation transfer. In: Sinanoglu O, editor. *Modern Quantum Chemistry*. New York: Academic Press. pp. 93–137.
- Gaits F, Hahn K. 2003. Shedding light on cell signaling: Interpretation of FRET biosensors. *Science's STKE* 165.
- Gratton E, Limkeman M, Lakowicz JR, Maliwal BP, Cherek H, Laczkó G. 1984. Resolution of mixtures of fluorophores using variable-frequency phase and modulation data. *Biophys J* 46:479–486.
- Hahn K, Touchkine A. 2002. Live-cell fluorescent biosensors for activated signaling proteins. *Curr Opin Cell Biol* 14:167.

- Hodgson L, Nalbant P, Shen F, Hahn K. 2006. Imaging and Photobleach Correction of Mero[hyphen (true graphic)] CBD, Sensor of Endogenous Cdc42 Activation. *Methods Enzymol* 406:140–156.
- Hodgson L, Pertz O, Hahn KM, Kevin FS. 2008. Design and optimization of genetically encoded fluorescent biosensors: GTPase Biosensors. *Methods Cell Biol* 85:63–81.
- Hodgson L, Shen F, Hahn K. 2010. Biosensors for characterizing the dynamics of rho family GTPases in living cells. . *Curr Protoc Cell Biol*. Chapter: 1–26.
- Ibraheem A, Campbell RE. 2010. Designs and applications of fluorescent protein-based biosensors. *Curr Opin Chem Biol* 14:30–36.
- Jameson D, Gratton E, Hall R. 1984. The measurement and analysis of heterogeneous emissions by multifrequency phase and modulation fluorometry. *Appl Spectrosc Rev* 20:55–106.
- Kraynov VS, Chamberlain C, Bokoch GM, Schwartz MA, Slabaugh S, Hahn KM. 2000. Localized Rac activation dynamics visualized in living cells. *Science* 290:333–337.
- Lakowicz JR, Masters BR. 2008. Principles of fluorescence spectroscopy, 3rd ed. *J Biomed Opt* 13:029901–029902.
- Lakowicz JR, Laczko G, Cherek H, Gratton E, Limkeman M. 1984. Analysis of fluorescence decay kinetics from variable-frequency phase shift and modulation data. *Biophys J* 46:463–477.
- Llères D, Swift S, Lamond AI. 2001. Detecting protein-protein interactions in vivo with FRET using multiphoton fluorescence lifetime imaging microscopy (FLIM) *Curr Protoc Cytom* 12:12.10.1–12.10.19.
- Machacek M, Hodgson L, Welch C, Elliott H, Pertz O, Nalbant P, Abell A, Johnson GL, Hahn KM, Danuser G. 2009. Coordination of Rho GTPase activities during cell protrusion. *Nature* 461:99–103.
- Parasassi T, De Stasio G, Ravagnan G, Rusch RM, Gratton E. 1991. Quantitation of lipid phases in phospholipid vesicles by the generalized polarization of Laurdan fluorescence. *Biophys J* 60:179–189.
- Periasamy A, Wallrabe H, Chen Y, Barroso M. 2008. Chapter 22: Quantitation of Protein-Protein Interactions: Confocal FRET Microscopy. *Methods Cell Biol* 89:569–598.
- Pertz O, Hahn KM. 2004. Designing biosensors for Rho family proteins-deciphering the dynamics of Rho family GTPase activation in living cells. *J Cell Sci* 117:1313–1318.
- Pertz O, Hodgson L, Klemke RL, Hahn KM. 2006. Spatiotemporal dynamics of RhoA activity in migrating cells. *Nature* 440:1069–1072.
- Suhling K, French PMW, Phillips D. 2005. Time-resolved fluorescence microscopy. *Photochem Photobiol Sci* 4:13–22.
- Tsien RY. 1998. The green fluorescent protein. *Annu Rev Biochem* 67:509–544.
- van Munster EB GT. 2005. Fluorescence lifetime imaging microscopy (FLIM). *Adv Biochem Eng Biotechnol* 95:143–175.
- Wallrabe H, Periasamy A. 2005. Imaging protein molecules using FRET and FLIM microscopy. *Curr Opin Biotechnol* 16:19–27.
- Yu W, So PT, French T, Gratton E. 1996. Fluorescence generalized polarization of cell membranes: A two-photon scanning microscopy approach. *Biophys J* 70:626–636.